



Assay Solution
A Solution Provider

Bovine SOD2/Mn-SOD ELISA Kit

Catalog #: AYQ-E10300 (96 wells)

User Manual

This kit is designed to quantitatively detect the levels of Bovine SOD2/Mn-SOD in cell lysates, serum/plasma and other suitable sample solution.

Manufactured and Distributed by:

AssaySolution

310 W Cummings Park, Woburn, MA, 01801, USA

Phone: (617) 238-1396, Fax: (617) 380-0053

Email: Info@assaysolution.com

FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC PURPOSES

Important notes

Before using this product, please read this manual carefully; after reading the subsequent contents of this manual, please note the following specially:

- The operation should be carried out in strict accordance with the provided instructions.
- Store the unused strips in a sealed foil bag at 2-8°C.
- Always avoid foaming when mixing or reconstituting protein solutions.
- Pipette reagents and samples into the center of each well, avoid bubbles.
- The samples should be transferred into the assay wells within 15 minutes of dilution.
- We recommend that all standards, testing samples are tested in duplicate.
- Using serial diluted sample is recommended for first test to get the best dilution factor.
- If the blue color develops too light after 15 minutes incubation with the substrate, it may be appropriate to extend the incubation time (Do not over-develop).
- Avoid cross-contamination by changing tips, using separate reservoirs for each reagent.
- Avoid using the suction head without extensive wash.
- Do not mix the reagents from different batches.
- Stop Solution should be added in the same order of the Substrate Solution.
- TMB developing agent is light-sensitive. Avoid prolonged exposure to the light.

ALWAYS REFER TO LOT SPECIFIC PROTOCOL PROVIDED WITH EACH KIT FOR INSTRUCTIONS.

Table of Contents

Contents	Page#
Intended Use	3
Assay Principle	4
Materials	4-5
Sample Preparation	5
Reagent Preparation	5-6
Assay Procedure	6-7
Result Calculation	
Typical data	7
Sensitivity	7
Spiking and Recovery	8
Reproducibility	8
Specificity	8
Sample dilution	8
Plate Layout	9
Troubleshooting Information	10-11

Catalog#	AYQ-E10300
Name	Bovine SOD2/Mn-SOD ELISA kit
Size	96 T
Storage	Store the unopened product at 2 - 8° C. Protect from light. Do not use past expiration date.
Gene ID	104973000
Gene Symbol	SOD2
Synonym	EC 1.15.1.1; indophenoxidase B; IPOB; IPO-B; manganese-containing superoxide dismutase; mangano-superoxide dismutase; Mn SOD; Mn superoxide dismutase; MnSOD; MVCD6; SOD, Mitochondrial; SOD2; superoxide dismutase [Mn], mitochondrial; superoxide dismutase 2, mitochondrial
Species	Bovine

Scientific Background

Superoxide dismutase 2, mitochondrial (SOD2), also known as manganese-dependent superoxide dismutase (MnSOD), is an enzyme which is encoded by the SOD2 gene on chromosome 6. A related pseudogene has been identified on chromosome 1. Alternative splicing of this gene results in multiple transcript variants. This gene is a member of the iron/manganese superoxide dismutase family. It encodes a mitochondrial protein that forms a homotetramer and binds one manganese ion per subunit. This protein binds to the superoxide byproducts of oxidative phosphorylation and converts them to hydrogen peroxide and diatomic oxygen. Mutations in this gene have been associated with idiopathic cardiomyopathy (IDC), premature aging, sporadic motor neuron disease, and cancer.

As a member of the iron/manganese superoxide dismutase family, this protein transforms toxic superoxide, a byproduct of the mitochondrial electron transport chain, into hydrogen peroxide and diatomic oxygen. This function allows SOD2 to clear mitochondrial reactive oxygen species (ROS) and, as a result, confer protection against cell death. As a result, this protein plays an antiapoptotic role against oxidative stress, ionizing radiation, and inflammatory cytokines.


Intended use

The kit is used to quantify the Bovine SOD2/Mn-SOD in serum/ plasma, cell culture supernatant and other suitable sample solution.

Standard range	156.25 – 10000 pg/mL
Assay time	4.5 hours
Validity	Six months
Store at	2-8 °C

Assay principle

The Bovine SOD2/Mn-SOD ELISA Kit is based on standard sandwich enzyme-linked immunosorbent assay technology. Anti-Bovine SOD2/Mn-SOD specific antibody has been pre-coated onto 96-well plate. Bovine SOD2/Mn-SOD present in the standards/samples bind to the capture antibody. Subsequently, biotinylated anti-Bovine SOD2/Mn-SOD detection antibody is added to form an Ab-Ag-Ab sandwich. After a washing step, streptavidin-HRP is added and the unbound conjugate is removed with wash buffer. Next, addition of HRP substrate, TMB, results in the production of a blue colored product that changes to yellow after the addition of acidic Stop Solution. The density of yellow color is directly proportional to the amount of Bovine SOD2/Mn-SOD captured on plate.



Materials supplied

1. Bovine SOD2/Mn-SOD standard:	10 ng/vial × 2
2. 96-well plate pre-coated with anti-Bovine SOD2/Mn-SOD Ab:	1
3. Sample Diluent buffer :	12 ml × 2
4. Detection antibody:	1 vial, dilution 1:60
5. Streptavidin-HRP:	1 vial dilution 1:40
6. Antibody Diluent buffer:	12 ml
7. Streptavidin HRP diluent buffer:	12 ml
8. Chromogen Solution A:	6 ml
9. Chromogen Solution B:	6 ml
10. Stop Solution:	6 ml
11. 20 × Wash Buffer:	25 ml
12. Plate sealers	2
13. Package insert	1

Materials required but not supplied

- 1x PBS.
- Standard plate reader capable of measuring absorbance at 450 nm.
- Adjustable pipettes and disposable pipette tips.
- Multi-channel pipettes, manifold dispenser or automated microplate washer.
- Distilled water.
- Absorbent paper.
- Materials used for sample preparation.

Sample Preparation and storage

Store samples to be assayed within 24 hours at 2-8°C. For long-term storage, aliquot and freeze samples at -20°C. Avoid repeated freeze-thaw cycles.

- Cell culture supernatant, tissue lysate or body fluids: Remove particulates by centrifugation, analyze immediately or aliquot and store at -20°C
- Serum: Allow the serum to clot in a serum separator tube (about 4hours) at room temperature. Centrifuge at approximately 1000 X g for 15 min. Analyze the serum immediately or aliquot and store frozen at -20°C.
- Plasma: Collect plasma using heparin as an anticoagulant. Centrifuge for 15 min at 1000 x g within 30 minutes of collection. Analyze immediately or aliquot and store frozen at -20°C. EDTA and citrate are not recommended as the anticoagulant.

Reagent Preparation

Standard

- Bovine SOD2/Mn-SOD: Standard solution should be prepared no more than 2 hours prior to the experiment. Two tubes of standard (10 ng /vial) are included in each kit. Use one tube for each experiment.
- Prepare 10000 pg/ml→156 pg/ml of Bovine SOD2/Mn-SOD standard solutions:
- Add 1.0 ml of Sample Diluent Buffer into one standard vial with 10 ng Bovine SOD2/Mn-SOD. Keep the tube at room temperature for 10 minutes and mix thoroughly. This is 10000 pg/ml standard solution.
- Label 6 Eppendorf tubes with 5000 pg/ml, 2500 pg/ml, 1250 pg/ml, 625 pg/ml, 313 pg/ml, 156 pg/ml respectively. Then make 2-fold serial dilution from 10000 pg/ml to 156 pg/ml in 1.5 ml tubes with sample diluent buffer.
- Make sure each tube has $\geq 250 \mu\text{l}$ of standard.

Note: *The standard solutions are best used within 2 hours.*

Preparation of detection anti-Bovine SOD2/Mn-SOD antibody working solution

- The stock solution is stable at 2-8 °C for up to 1 month. After opening the vial use within 1 month. For long-term storage, please aliquot and store at -20 °C. Avoid freeze-thaw cycles.
- The working solution should be prepared no more than 2 hours prior to the experiment

- The reagent is supplied as 60X concentrate. Empty the total contents into 10.62 ml of Antibody Diluent Buffer or prepare the solution separately in a volume as needed. The solution should be mixed thoroughly.
- The total volume should be: 0.1 ml/well x the number of wells (Allowing 0.1-0.2 ml more than total volume).

Preparation of Streptavidin-HRP working solution

- The solution should be prepared no more than 1 hour prior to the experiment.
- The total volume should be: 0.1 ml/well x the number of wells (allowing 0.1-0.2 ml more than total volume).
- Streptavidin-HRP should be diluted 1:40 with Streptavidin-HRP Diluent buffer and mixed thoroughly.

Preparation of TMB Substrate solution

- The solution should be prepared no more than 10 min prior to the experiment.
- The total volume should be: 0.1 ml/well x the number of wells (allowing 0.1-0.2 ml more than total volume).
- Mix equal volumes of chromogen solutions A and B. Protect from light.

Wash Buffer

- If crystals have formed in the 20X wash buffer, warm to room temperature and mix gently until the crystals have completely dissolved.
- Dilute 25 ml Wash Buffer Concentrate (20X) to a total volume of 500 ml with distilled water.

Assay Procedure

Bring all reagents to room temperature before use. Bovine SOD2/Mn-SOD standard curve should be prepared for each experiment. The user will decide sample dilution factor by rough estimation of Bovine SOD2/Mn-SOD concentration in samples.

1. Add 100 μ l of sample or standards per well. Add 0.1 ml of the sample diluent into the control well (Zero well). Cover with an adhesive strip and incubate at room temperature for 2 hours.
Note: We recommend that each Bovine SOD2/Mn-SOD standard solution and each sample is measured in duplicate.
2. Aspirate each well and wash with Wash Buffer, repeating the process two times for a total of three washes. Wash by filling each well with Wash Buffer (300 μ l) using a squirt bottle, manifold dispenser, or auto-washer. Complete removal of liquid at each step is essential for good performance. After the last wash, remove any remaining Wash Buffer by aspirating or by inverting the plate and blotting it against clean paper towels.
3. Add 100 μ l of the Detection Antibody working solution to each well. Cover with a new adhesive strip and incubate at room temperature for 2 hours.
4. Repeat the aspiration/wash as in step 2.
5. Add 100 μ l of the working solution of Streptavidin-HRP to each well. Cover the plate and incubate at room temperature for 30 min. Avoid placing the plate in direct light.
6. Repeat the aspiration/wash as in step 2 for three times.

7. Add 100 μ l of TMB substrate solution to each well. Cover and incubate at room temperature for 5 -15 min until a gradient develops and you see visible color in the 2nd lowest concentration well. Protect from light. Do not over-develop.
8. Add 50 μ l Stop Solution to each well. Mix well.
9. Read the Optical Density (O.D.) at 450 nm using a microtiter plate reader immediately.


Result calculation

For calculation, (the relative O.D.450) = (the O.D.450 of each well) – (the O.D.450 of Zero well). The standard curve can be plotted as the relative O.D.450 of each standard solution (Y) vs. the respective concentration of the standard solution (X). The Bovine SOD2/Mn-SOD concentration of the samples can be interpolated from the standard curve.

Note: if the samples measured were diluted, multiply the dilution factor to the concentrations from interpolation to obtain the concentration before dilution

Typical data

This standard curve was generated at the AssaySolution laboratory for demonstration purpose only. A standard curve must be run with each assay.



Sensitivity

The sensitivity or minimum detectable dose (MDD) of Bovine SOD2/Mn-SOD was determined to be 125 pg/ml. MDD is defined as the Bovine SOD2/Mn-SOD concentration resulting in an O.D.₄₅₀ value that is 2 standard deviations higher than blank.

Spiking and Recovery

Recovery was determined by spiking the following matrices with various concentrations of Bovine SOD2/Mn-SOD.

Sample Type	Average Recovery (%)	Range (%)
Cell lysates	94.7	89-101
Serum	102.4	95-109

Reproducibility

- Inter-assay- <10.7%
- Intra-assay- <8.8%

Specificity

This kit recognizes both natural and recombinant Bovine SOD2/Mn-SOD.

Sample Dilution

The user may need to determine the dilution factor in a preliminary experiment. If required, samples should be diluted in sample diluent buffer.

*For additional troubleshooting information please
email us at info@assaysolution.com*

Plate Layout

1								
2								
3								
4								
5								
6								
7								
8								
9								
10								
11								
12								
A	B	C	D	E	F	G	H	

Troubleshooting Information

High Background

Probable Cause:	Solution/ Action
High incubation temperature:	Incubate at room temperature (25 °C) throughout the procedure
Insufficient washing of the plate:	Fill the wells with wash buffer and aspirate completely for the next wash Increase the number of washes Add soak time (20-30 seconds) in between the washes Use automated plate washer, if available and check that all the channels are operating properly
Concentrated streptavidin-HRP	Streptavidin-HRP was not diluted properly Dilute the streptavidin-HRP as mentioned in the manual
Light exposure during substrate incubation	The TMB substrate is light sensitive and turns to blue color in the presence of light. The incubation must be carried out in dark.
Stop solution not added	Color will continue to develop if stop solution is not added
Diluents came with the kit were not used	Standards/ sample, detection antibody and streptavidin-HRP must be diluted in the respective buffers came with the kit. Do not use buffers from other kits
Contaminated solutions	Prepare fresh working solutions

Poor Standard Curve

Probable Cause:	Solution/ Action
Improper standard reconstitution:	Spin the vial briefly before opening Reconstitute the standard as mentioned in the manual. After reconstitution, leave it atleast for 10 minutes at room temperature Do not store and reuse diluted standards
Curve fitting problem:	Log transform the values on both axes Use 4-PL/ 5-PL curve fitting programs
Incubation temperature/ time	Use the recommended standard incubation conditions
Poor dilutions	Pipetting error. Check pipetting technique and calculations. Use calibrated pipettes.

No Signal

Probable Cause:	Solution/ Action
Omission of reagent(s):	Read the manual entirely. Check that all the reagents are added in the correct order as stated in the manual
Incorrect detection antibody was used:	Use the detection antibody came with the kit
Chromogen solutions were mixed improperly	Use the recommended procedure to prepare the TMB substrate
HRP inhibitor in sample/ buffers	Check that the samples/ buffers do not have sodium azide as it will inhibit peroxidase reaction.
Vigorous washing	If the washing is done manually, pipette the wash buffer gently.
Dried wells	Do not allow the wells to dry out during the assay. Seal with the supplied adhesive cover during incubations
Improper plate reader settings	Check the wavelength and read the plate again

Erratic duplicate OD values

Probable Cause:	Solution/ Action
Insufficient washing of the plate	Fill the wells with wash buffer and aspirate completely for the next wash Increase number of washes Add soak time (20-30 seconds) in between the washes Use automated plate washer, is available and check that all the channels are functioning properly
Poor dilutions	Pipetting error. Check pipetting technique and calculations. Use calibrated pipettes.
Improper mixing of samples/ buffers	Mix the samples well before pipetting Thoroughly mix the working solutions of detection antibody/ streptavidin-HRP
Contamination from other wells	Do not reuse the adhesive covers from previous assay setups Change pipette tips during reagent addition. If same pipette tip is being used to dispense reagents, care should be taken, not to touch the solution in the well
Precipitates in the samples/ buffer	If precipitates are visible in wash buffer concentrate, keep it at 37 °C for 10-15 minutes until no precipitates are visible Centrifuge the samples to remove particulate matter
Dried wells	Do not allow the wells to dry out during the assay. Seal with the supplied adhesive cover during incubations